

Table 1. Reporting Limit for each method [ng/L, nanogram per liter; NA, not analyzed; NC, not calculated]

Compound	CLAM <sup>1</sup>	Discrete	POCIS
	ng/L	ng/L	ng/L
Caffeine	9	60	22
Ethyl citrate	4	160	8
Cotinine	28	800	64
Dichlorvos	13	NA	210
Phenol	9	160	26
Tri(2-chloroethyl) phosphate (Fyrol CEF)	13	100	29
Acetophenone	7	400	54
Bromacil	14	360	130
Metalaxyl	18	120	170
5-Methyl-1H-benzotriazole	182	1200	NC
Bromoform	19	100	91
Indole	4	80	26
<i>para</i> -Cresol	7	80	26
Isoquinoline	7	160	27
<i>N,N</i> -diethyl- <i>meta</i> -toluamide (DEET)	3	60	12
Carbaryl	7	160	65
3-Methylindole (Skatol)	4	36	39
Methyl salicylate	8	44	56
Isophorone	4	32	8
Atrazine	13	NA	39
Diethyl phthalate	3	NA	36
Isoborneol	8	80	NC
Tetrachloroethylene	27	120	100
Tri(2-butoxyethyl) phosphate	100	800	320
Camphor	8	44	65
Benzophenone	6	80	66
Naphthalene	2	40	14
Carbazole	6	30	15
Metolachlor	3	28	5
1,4-Dichlorobenzene	9	40	230
Anthraquinone	3	160	33
Menthol	34	320	280
Isopropyl benzene (Cumene)	4	300	30
3- <i>tert</i> -Butyl-4-hydroxyanisole (BHA)	16	600	NC
Prometon	20	120	19
Bisphenol A	12	NA	22
Tri(dichloroisopropyl) phosphate (Fyrol FR2)	24	160	35
1-Methylnaphthalene	4	22	30
2-Methylnaphthalene	4	36	30
Tributyl phosphate	3	160	15
Diazinon	28	160	42
3,4-Dichlorophenyl isocyanate	97	NA	45
4-Cumylphenol	3	60	NC
4- <i>tert</i> -Octylphenol	1	140	13
4-Nonylphenol monoethoxylate (NP1EO)	19	NA	190
4-Nonylphenol diethoxylate (NP2EO)	38	5000	320
2,6-Dimethylnaphthalene	4	60	32
Anthracene	2	10	16
Phenanthrene	2	16	16
<i>para</i> -Nonylphenol-Total	25	2000	140
4-Octylphenol diethoxylate (OP2EO)	4	1000	32
Chlorpyrifos	29	160	23
Triclosan	28	200	290
Triphenyl phosphate	6	120	88
Pentachlorophenol	321	NA	500
d-Limonene	18	80	130
4-Octylphenol monoethoxylate (OP1EO)	3	1000	31
Fluoranthene	2	24	18
Pyrene	2	42	18
Hexahydrohexamethylcyclopentabenzopyrane (Galaxolide; HHCB)	3	52	10
4-n-Octylphenol	2	60	NC
Tonalide (AHTN)	3	28	10
Benzo[a]pyrene	1	60	19
2,2',4,4'-Tetrabromodiphenylether (PBDE 47)	3	NA	NC
Diethylhexyl phthalate	2	NA	54
Cholesterol	24	2000	68
3- <i>beta</i> -Coprostanol	25	1800	68
<i>beta</i> -Sitosterol	45	4000	140
<i>beta</i> -Stigmastanol	22	2600	68

<sup>1</sup> CLAM MRL calculated using the maximum volume of all of the samples, 67.5 liters.